



**Lady Reading Hospital (LRH)
Medical Teaching Institution (MTI)**

Peshawar
Contact No: 091-9211927

**BID SOLICITATION DOCUMENTS
For
Medical Non-Medical Tender & Accessories
For the Year 2024-2025**

Note: The prospective bidder is expected to examine the Bidding Documents carefully, including all Instructions, Terms & Conditions, and Specifications etc. Failure to furnish all information required by the Bidding documents or submission of a Bid not substantially responsive to the Bidding Documents in every respect would result in the rejection of the Bid.

Asst Manager Material Management

Bio Medical Engineer

A.D Legal

Director Finance

Nursing Director

Associate Hospital Director

Medical Director

Hospital Director

TABLE OF CONTENTS

S #	CONTENTS	PAGE #
1.	Invitation For Bids	02
2.	Instruction to Bidders	03
3.	Bid Security	04
4.	Bid Validity	04
5.	Blacklistment of defaulted bidder/contractor	04
6.	Dispute and Controversies/Dispute Resolution	04
7.	List of Items	05-32
8.	Evaluation Criteria	33
9.	Award of Contract	34
10.	Payment	34
11.	Tender Notice	35

INTRODUCTION:

Lady Reading Hospital/ Medical Teaching Institution (MTI) Peshawar invites sealed bids from the eligible bidders for procurement of “**Medical Non-Medical & Accessories Tender FY 2024-25**” for Hospital under Open Competitive Bidding under rule 6(2) (b) “**Single Stage Two Envelope**” bidding procedures of Khyber Pakhtunkhwa Public Procurement Regulatory Authority (KPPRA) Rules 2014.

1) INSTRUCTIONS TO BIDDERS:

1. This Bidding procedure will be conducted in light of Khyber Pakhtunkhwa Public Procurement Regulatory Authority (KPPRA) Laws, Rules made there under, along with Standard Bidding Documents.
2. The bids will be opened on _____ at _____ AM in presence of the bidders/representatives who choose to attend.
3. Pre-bid meeting with the interested bidders will be held on _____-2024 at _____:00 AM in AHD office /Committee room of AHD of the Institution.
4. Any bid received after the deadline for submission of bids shall not be entertained and shall be returned unopened to the Bidder.
5. The bid should be complete in all respect and must be signed by the bidder.
6. All prices quoted must be in Pak Rupees (PKR) and should include all applicable taxes. If not specifically mentioned in the Quotation, it will be presumed that the prices include all the taxes.
7. Bidders are essentially required to provide correct and latest postal/email/web addresses, phone/mobile/fax numbers for actively and timely communication.
8. For any query, clarification regarding Services / Bid Solicitation Documents, the applicants may send a written request at least one day prior to the opening date.
9. The Bidder may after its submission withdraw its bid prior to the expiry of the deadline prescribed for submission of bids. Withdrawn bids will be returned unopened to the Bidders.
10. Any bid not received as per terms and conditions laid down in this document are liable to be ignored. No offer shall be considered if:
 - a. Received without earnest money;
 - b. It is received after the date and time fixed for its receipt;
 - c. The offer is ambiguous;
 - d. The offer is conditional i.e. advance payment, or currency fluctuations etc.;
 - e. The offer is from blacklisted firm in any Federal / Provincial Govt. Deptt:
 - f. Only typed tender on original prescribing letter pad, sealed & signed should be submitted, the quoted Price must be pre printed and hand written quoted price will not be acceptable. The tenders must be according to hospital specification; alternate rates (**Double rates for single Items**) will not be accepted.
11. Usage of correction fluid & corrections are strictly prohibited unless duly initialed.

12. Any erasing / cutting etc. appearing on the offer, must be properly signed by the person signing the tender.
13. Bids will be rejected if the Bid is in some way connected with bids submitted under names different from his own.
14. In case of Bid Tie, the decision will be taken by making toss/draw/Recall sealed quotation in front of the bidders.
15. Any direct or indirect effort by a bidding firm to influence this institution during the process of selection of a bidder or award of contract may besides rejection of its bid result into its disqualification from participation in the LRH Peshawar's future bids.

2) ELIGIBILITY CRITERIA:

- Bidders must give compliance to the below mentioned clauses as these are mandatory to being eligible for the bidding process. Relevant certificates must be attached.
- The bidder shall provide an undertaking that the bidder has not been declared black listed by any Governmental/ Semi-Governmental institutions.
- Bidders shall not be eligible to bid if they are under a declaration of Ineligibility for corrupt and fraudulent practices issued by any government organization in accordance with the Section 44(1) KPP Rules 2014.

3. GENERAL CONDITIONS: -

1. LRH Peshawar shall evaluate the proposal in a manner prescribed in advance, without reference to the price and reject any proposal which does not conform to the specified requirements.
2. Alternative bid shall not be considered and shall be rejected by the Competent Authority.
3. At any time prior to the deadline for submission of bids, LRH Peshawar may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective Bidder, modify the bidding documents by amendment.
4. If a bid is not substantially responsive, it will be rejected by the Procuring Entity and may not subsequently be made responsive by the Bidder by correction of the nonconformity.
5. LRH MTI Peshawar may accept or reject any or all of the bids under Rule 47 of KPPRA Rules, 2014.

**Lady Reading Hospital
Medical Teaching institute
Peshawar
Phone: 091-9211927**

INVITATION FOR BIDS

Hospital Director, Medical Teaching Institute, **Lady Reading Hospital Peshawar** Invites sealed tenders under National Competitive Bidding for the procurement of “**Medical Non-Medical & Accessories FY 2024-25**” for Hospital, under rule 6(2) (b) “*single stage Two envelope procedure*” of KPPRA Rules 2014, from Income / Sales tax, reflected on Active Taxpayer List of FBR.

The bidders are required to submit bid security **Rs-200,000/-** in the name of Hospital Director LRH Peshawar. Pre-bid meeting with the interested bidders will be held on **-2024 at 11:00 AM** at AHD/Committee room of Associate Hospital Director of the institution.

The tenders complete in all respect must reach the undersigned by **11:00 AM on /2024**, which will be opened at 11:30 AM on the same day in conference room / AHD office of the Hospital in the presence of the procurement committee and the bidders / representatives who may choose to attend.

Competent Authority reserves the right to reject any or all the bids as per provisions contained in Rule 47 of KPPRA Rules 2014.

4. BID Security

Bid security **Rs-200,000/-** in favor of “Hospital Director LRH Hospital”
The Bid Security in shape of Pay Order is not accepted

The bid security may be forfeited:

- i) If a Bidder withdraws its bid during the period of bid validity specified by the Bidder on the Bid Form; or
- ii) In the case of a successful Bidder, if the Bidder fails to sign the contract
- iii) If a successful Bidder unable to supply the items after issuance of 3 reminders, order will be placed to the next lowest bidder.

5. BID VALIDITY:

- i) The bids should be valid **for 180 days**.
- ii) In exceptional circumstances, LRH Hospital may solicit the Bidder’s consent to an extension of the period of validity reasons shall be recorded in writing. The request and the responses there to shall be made in writing. The bid security provided shall also be suitably extended. A Bidder may refuse the request without forfeiting its bid security. A Bidder granting the request will not be required nor permitted to modify its bid, except as provided in the bidding document.

Medical Non-Medical & Accessories Tender
FY 2024-25

S.no	Equipment Name	Hospital Specification	Firm Specification	Pack Size	Unite Price	Pack Price
1	ECG Clips for ECG Machine	Adult/ Pediatrics				
2	ECG Clip Electrode					
3	EEG Disk Electrode					
4	ECG Cable Compatible for Cardiomate (Machine)	10 leads				
5	ECG Cable Compatible for Mindray IMEC-12 cardiac monitor	Imported				
6	Burner	For local autoclave				
7	NIBP Simulator (FLUKE)	FLUKE				
8	ECG Simulator (FLUKE)	FLUKE				
9	Defibrillator analyzer (FLUKE)	FLUKE				
10	Battery for Defibrillator	FLUKE				
11	Spo2 Probe Phillips Monitor					
12	Spo2 analyzer(FLUKE)	FLUKE				
13	Electrosurgical (ESU) Analyzers (FLUKE)	FLUKE				
14	Infusion analyzer (FLUKE)	FLUKE				
15	Radiometer for incubator and warmer (FLUKE)	FLUKE				
16	XDCR FLOW 0-200 L/MIN Flow Transducer for GE CARSCAPE					
17	Battery for ACOMA ventilator					
18	Pressure pipe for anesthesia machine (per feet)					
19	Battery for infusion Pump					
20	Battery for ventilator					
21	Expiratory Flow sensor for ventilators Oricare					
22	Inspiratory Flow sensor for ventilators Oricare					
23	Inlet Filter ventilators for Oricare					
24	Battery for ventilators Oricare					
25	Compressor Motor Pump for ventilators Oricare					

26	Compressor PCB Board for ventilators Oricare					
27	Power Supply for ventilators Oricare					
28	Mixer Tank for ventilators Oricare					
29	Relief valve for ventilators Oricare					
30	Inspiratory Diaphragm for ventilators Oricare					
31	One-way diaphragm for ventilators Oricare					
32	Power Board Perlong Anesthesia Machine Parts					
33	Soda line Jar for Perlong Anesthesia Machine Parts					
34	Vaporizer for anesthesia (Perlon) UK					
35	Alarm Value for Perlong Anesthesia Machine Parts					
36	Pressure Value for Perlong Anesthesia Machine Parts					
37	Switch for Perlong Anesthesia Machine Parts					
38	One-way check value for Perlong Anesthesia Machine Parts					
39	Chlorhexidine solution spray					
40	Boiler Instruments					
41	Server Rack 42U with PDU's					
42	Aprons Lead 0.5mm (Imported)					
43	Vacuum Pump					
44	Ultra-sonic nebulizer					
45	Analogue weight scale with height measurement					
46	Oxygen concentrator 5LTR	-				
47	Laryngoscope Adult	-				
48	BIPAP Machine (Philips)	-				
49	CPAP Machine (Philips)	-				
50	Syringe pump (USA/UK/JAPAN EU)	Syringe Pump for Fluid Administration.				
		Rate Mode, Time Mode, Body Weight Mode, Micro-infusion mode.				
		Flow Rates:				

	0.01-999.9 ml/hr				
	Display:				
	2.4”TFT,240x320 Pixels,262K colors, Digital Color Display of set parameters and viewing angle: all 80°				
	Universal syringe acceptance capability for Disposable, Plastic available.				
	Size: 2, 3, 5, 10, 20,30, 50, 60 ml.				
	Moisture Protection:				
	IP34(water splashing from any direction)				
	Fast Startup				
	Colour Display				
	Intuitive handling				
	Drive Accuracy:				
	± 2% in compliance with IEC/EN 60601-2-24				
	Automatic and Manual Bolus with 0.10-1800ml/hr.				
	Occlusion Pressure:				

	9 levels.				
	Drug Library:				
	3,000 drugs including all parameters in up to 30 drug categories.				
	Dose limit and dose error reduction system available.				
	Display of Delivery Rate, Current Infusion Status, VTBI, Total Volume, Syringe Brand, Pressure Limit, Battery Capacity, Infusion Volume, Drug Name, Remining Time, Alarms.				
	Automatic adoption of controls according to syringe set available.				
	Quick freed/rapid infusion facility available.				
	Rechargeable battery and mains operated 220V,50Hz.				
	Saftey Audible and acoustic alarms for :				
	Occlusion Pressure				
	End of infusion				

	Low Battery				
	Inproper syringe insertion				
	Technical alarms etc				
	Battery Backup:				
	Rechargeable maintenance free dry Battery with 10hrs at the rate of 25ml/hr.				
	Stakable:				
	stakable with each other upto 3 pumps.				
	Compatibility with Docking Station:				
	compatible with docking station.				
	Dose Rate Calculation:				
	Calculate the delivery rate on the basis of the following parameters.				
	Concentration value (mg, µg, IE or mmol)				
	Dosage value				
	Dosage unit				
	Time (mg per kg/min; mg/kg/h; mg/kg/24h)				
	Optional: Patient Weight				
	Bolus Application in mg, µg, mmol				

51	standard Oscope	welch Allyn/Riester/keeler/Hiene OR equivalent				
52	Infusion Pump (USA/UK/JAPAN EU)	Infusion Pump for Fluid Administration.				
		Simple, Safe, and Robust Design.				
		Integrated Dose Guard drug library.				
		Intuitive user interface due to self-explanatory menu.				
		Integrated pole clamp and handle.				
		Easy to use disposables with color coded connectors.				
		Guiding interface to prevent operating errors.				
		Predefined infusion rates for patient safety - B. Braun Dose Guard ®.				
		Flow Rates:				
		0.1-999.9 ml/hr				
Display:						

		2.4”TFT,240x320 Pixels,262K colors, Digital Color Display of set parameters and viewing angle: all 80°				
		High resolution colour display.				
		Visibility from all angles.				
		Colour coded drug library.				
		Universal pressure Infusion IV set Compatibe.				
		Drive Accuracy:				
		± 5% in compliance with IEC/EN 60601-2-24				
		Display of Delivery Rate,Current Infusion Status,Volume to be infused, Pressure Limit,Battery Capacity,Infusion Volume,Drug Name,Remining Time,Alarms.				
		Automatic adoptation of controls according to infusion set.				
		Quick freed/rapid infusion facility.				

	Rechargeable battery and mains operated 220V,50/60Hz.				
	Safety Audible alarms for :				
	Occlusion Pressure				
	End of infusion				
	Low Battery				
	Inproper syringe insertion				
	Technical alarms etc				
	Battery Backup:				
	Battery backup 8 hrs at the rate of 25ml/hr.				
	Moisture Protection:				
	IP34(water splashing from any direction)				
	Bolus Option:				
	Automatic and Manual Bolus with 1-1200ml/hr.				
	Manual bolus				
	Bolus with volume preselection				
	Bolus with preselection of volume and bolus duration				
	Occlusion Pressure:				
	9 levels.				
	Stakable:				

	stakable with each other upto 3 pumps.				
	Drug Library:				
	Customizable 3,000 drugs including all parameters in up to 30 drug categories.				
	Docking Station Compatible upto 18 pumps.				
	Fast Startup				
	Colour Display				
	Intutive handling				
	Dose Rate Calculation:				
	Calculate the delivery rate on the basis of the following parameters.				
	Concentration value (mg, µg, IE or mmol)				
	Dosage value				
	Dosage unit				
	Time (mg per kg/min; mg/kg/h; mg/kg/24h)				
	Optional: Patient Weight				
	Bolus Application in mg, µg, mmol				
	Accessories:				
	Built in Universal Pole Clamp, Built in Rechargeable Battery.				

53	Standard Ophthalmoscope	(welch Allyn/Rister/Keeler/Heine)				
54	IGBT	-				
55	Centrifuge Machine heavy duty					
57	Intermittent Pneumatic Compression Pump					
58	Cardiac prob for U/S machine					
59	Mosquito Light					
60	Chest percuser machine	imported				
61	Vein view finder					
62	Xray Viewer Double LED	-				
63	Centrifuge Machine (High Through Put)	Bench Top Centrifuge for serum/ Plasma separation				
		500-6000 RPM				
		Capacity: 24 tubs hang out type				
		Built- in automatic braking				
		At least 5 years with parts and services				
64	Micro – Centrifuge Machine For PCR	Runtime Programmable 30 Sec-60 Min				
		Max RCF: Programmable approximately Max 15000-22000				
		Capacity: 24 x 1.5/2ml				
		Display: Digital Display				
		Power supply: 220 Volt				

65	Oxygen concentrator 05 LTR Dual Outlet	Oxygen concentrator to provide oxygen from ambient air.				
		Oxygen concentration should be measured at the flow meter by oxygen sensing device (OSD)				
		Should have oxygen purity indicator/alarm for low oxygen concentration.				
		Sound level ≤ 55 dB				
		Superior grade of molecular sieve bed of minimum working hour of 15,000hrs.				
		Integrated Timer should be there.				
		Maintenance free rotary poppet valve.				
		Should have audio and visual alarm for Low oxygen concentration, low pressure and power failure.				
		The unit should be movable and should be provided with four wheels with at least two breaks.				

		Oxygen concentration 90% ±3% at 10LPM flow.				
		Oxygen output, approx.: 10 LPM.				
		Oxygen outlet Pressure: Minimum 10 psi.				
		Double outlet for oxygen Delivery controlled from two separate pressure compensated flow meter to ensure accurate flow display.				
		Masks: Necessary mask suitable to Neonatal should be provided: 2 sets.				
		Warranty 1 year				
66	Oxygen concentrator 10 LTR Dual Outlet	Oxygen concentrator to provide oxygen from ambient air.				
		Oxygen concentration should be measured at the flow meter by oxygen sensing device (OSD)				

	Should have oxygen purity indicator/alarm for low oxygen concentration.				
	Sound level ≤ 55 dB				
	Superior grade of molecular sieve bed of minimum working hour of 15,000hrs.				
	Integrated Timer should be there.				
	Maintenance free rotary poppet valve.				
	Should have audio and visual alarm for Low oxygen concentration, low pressure and power failure.				
	The unit should be movable and should be provided with four wheels with at least two breaks.				
	Oxygen concentration 90% $\pm 3\%$ at 15LPM flow.				
	Oxygen output, approx.: 15 LPM.				
	Oxygen outlet Pressure: Minimum 15 psi.				

		Double outlet for oxygen Delivery controlled from two separate pressure compensated flow meter to ensure accurate flow display.				
		Masks: Necessary mask suitable to Neonatal should be provided: 2 sets.				
		Warranty 1 year				
67	SPO2 Probe for Mindray Patient Monitor	Spo2 Probe Mindray patient monitor				
68	Anesthesia Machine Sensor	Anesthesia Machine Sensor for Mindray				
69	DC Power supply 10 Volt	10 Volt				
70	DC Power supply 12 Volt 20AMP	12 Volt				
71	DC Power supply 15 Volt 15AMP	15 Volt				
72	Fluid Warmer	Input: 230v 50 Hz				
		Set Temperature $+42C \pm 3C$ at $+ 22c$ ambient temperature				
		Capacity: 270 Liters				
		Safety Class: 1 Protective earthing				
		Or Updated Model				

73	Spo2 Probe for CM-500 Patient Monitor	Spo2 Probe for CM-500 Patient Monitor				
74	NIBP Cuff with Hose Line for OPERON	NIBP Cuff with Hose Line for OPERON				
75	NIBP Cuff with Hose Line for Pro M12E	NIBP Cuff with Hose Line for Pro M12E				
76	NIBP Cuff with Hose Line for Mindray	NIBP Cuff with Hose Line for Mindray				
77	Battery for Patient Monitor	Battery for Patient Monitor				
78	DC Power Supply 24Volt, 10A	DC Power Supply 24Volt, 10A				
79	220 Volt to 110volt supply or Transformer	220 Volt to 110volt supply or Transformer				
80	Suction Machine	It should have maintenance free piston/cylinder technology with stand be heat resistant with high tech material for long lasting.				
		It should have two autoclavable PSU (Polysulfone) Jars of 3/5 Liters with change switch from are Jar to another.				
		It Should be noiseless and vibration free.				

		It should have high flow of at least sixty liters per minute with -95 KPA (-713 mmhg or more)				
		It should have provision of variable flow rate of 40L/50L/60L.				
		It should have a foot on/off switch for hands free operation.				
		It should be made with high quality housing material.				
		Should be CE and FDA Approved.				
		Warranty 01 Year				
81	Auto Calve Machine (ISO certified)	Dimension: 16"x30" or more				
		Standard Features				
		Gas /Electrical				
		12 Gauge Inner Sheet				
		Working Pressure 25LB				
		Heavy Duty				
82	BP Cuff for Patient Monitor Welch Allyn Model CSM	BP Cuff for Patient Monitor Welch Allyn Model CSM				

83	SPO2 Probe for Patient Monitor Welch Allyn Model CSM	SPO2 Probe for Patient Monitor Welch Allyn Model CSM				
84	Back up Battery for Patient Monitor Schiller model Truescope classic	Back up Battery for Patient Monitor Schiller model Truescope classic				
85	Spo2 probe for Patient Monitor PMED Model PM-E12	Spo2 probe for Patient Monitor PMED Model PM- E12				
86	Leads for Holter Monitor (EDAN SE-2003)	Leads for Holter Monitor (EDAN SE-2003)				
87	BP Cuff for Ambulatory BP Monitor (Adult/Peads)	BP Cuff for Ambulatory BP Monitor (Adult/Peads)				
88	BP Cuff Bladder with line for Fresenius Dialysis Machine					
89	Weight Machine for Wheel Chair	With support				
90	Tube light for Infant Phototherapy	20 Watt				
91	LCD Digital Temperature and Humidity meter Indoor					
92	LCD Digital Temperature and Humidity meter Indoor/Outdoor					
93	Neonatal CPAP Machine	Bubble CPAP machine with compressor (FDA/CE/ISO13485)				
		• CPAP generator with pressure range from 3 to 10 cm of water.				
		• Capable of giving nasal/nasopharyngeal CPAP.				

	<ul style="list-style-type: none"> • Integrated pneumatic Air and oxygen blender calibrated with flow from 0-15 lit/min. 				
	<ul style="list-style-type: none"> • Safety mechanism for relief of excessive pressure through pressure relief valve/regulator. 				
	<ul style="list-style-type: none"> • Soft nasal prongs. 				
	<ul style="list-style-type: none"> • Alarms for the device – Low/high Temperature Flow increase/decrease alarm O2 pressure low alarm Air pressure low alarm 				
	<ul style="list-style-type: none"> • Flow meters: O2 with each piece. 				
	<ul style="list-style-type: none"> • Power 220-230 volts 50 Hz. 				
	<ul style="list-style-type: none"> • Power pack (UPS with battery back of minimum 1 hour). 				
	<ul style="list-style-type: none"> • System should be quoted with pole assembly to incorporate the whole CPAP machine. 				

		<ul style="list-style-type: none"> • Standard accessories with each equipment Heated wire servo-controlled humidifier 				
		5 ml test lung				
		Disposable patient circuits 30 Disposable nasal prongs 30 (10 each of different neonatal sizes)				
		Warranty: For 2 years for all parts excluding consumables.				
94	Oil filter for compressor					
95	Air filter cartridge					
96	Compressor Oil					
97	Relay for RO					
98	UV tube for RO					
99	Vein Detector Device	CE Approved				
100	Membrane	Main RO/ICU RO				
101	Membrane	Portable RO				
102	Softener filter vessel	Main RO/ICU RO				
103	Carbon filter/vessel	Main RO/ICU RO				
104	Multimedia filter/vessel	Main RO/ICU RO				
105	Magnetic contactor	Main RO/ICU RO				
106	Feeding/storage motor pump	Main RO/ICU RO				
107	RO fitting (Run Per Feet)	Main RO/ICU RO				
108	Complete RO service (Main RO/ICU RO)					
109	Pressure switch for RO					

110	Phototherapy LED panel	For nursery phototherapy lights				
111	Incubator motor	For baby incubator				
112	Incubator power supply	For baby incubator				
113	9-volt battery					
114	Relay 48-volt Dc	For baby incubator/warmer				
115	Heating glass rod 500watt	For baby warmer				
116	Battery for CX-100 battery	For CX-100 pulse Oximeter				
117	Infusion pump battery	BYS-820/SK-600I/				
118	Heating element 5000watt with seal	For local Autoclave				
119	Butter fly nut	For local Autoclave				
120	Water level column with seal	For local Autoclave				
121	Gas kit/door seal	For local Autoclave				
122	Autoclave steam tank	For local Autoclave				
123	Safety Valve	For local Autoclave				
124	Teflon tape					
125	Water inlet valve	For local Autoclave				
126	Steam outlet valve	For local Autoclave				
127	Battery for Orthopedic Drill					
128	LED GLASS					
129	Disposable Deflator Pads					
130	ETT Holder					
131	Head ring for Pronin Patients					
132	High Flow nasal Canula					
133	Disposable Filter for Vacuum Regulators					
134	Delivery Cushion					

135	Electric Autopsy Saw with Standard Blades					
136	Biposy forceps for Branchoscopy					
137	Biposy forceps for Thoracoscopy					
138	Spo2 Probe for Data Scope					
139	Original Bulb for NIDE Slit Lamps					
140	Thyraode Shield					
141	ECG Lead wire Set					
142	LED GLASS					
143	OSRAM Bulb 15v/150w					
144	Butterfly Valve 3/8					
145	Nasoendoscopes Flexible					
146	Vacuum Cup rubber					
147	Spo2 Lead for GE Healthcare Monitor					
148	ECG Lead for Mindray					
149	ECG Lead for NH					
150	Sterilization Paper (All Sizes)					
151	Liquid Level Sensor					
152	Flate roll (All Sizes)					
153	Gusseted roll (All Sizes)					
154	Crepe Paper (All sizes)					
155	Iv stand (SS)					
156	CAUTERY MACHINE					
157	ANESTHESIA MONITOR STAND					
158	SONICAIDS					
159	CTG MACHINE					
160	BULBS FOR LARYNGOSCOPE BLADES					
161	UV Lamp with Glass for RO Plant					
162	BULB 12VOLTS 100WATT					
163	NIBP CUFF CARDIAC MONETER					
164	WATER TRAP CO2 BENEVIEW T6					
165	COLLIMATOR LIGHT (12 V, 150 W)					
166	SPO2 PROBE PEADS PHILIPS					
167	NASOENDOSCOPES, FLEXIBLE & RIGID					
168	VACCUM CUP RUBBER (LARGE)					
169	OXYGEN SENSOR /CELLS FOR CARESCAPE VENTILATOR					
170	SPO2 PROBE FOR NEW TECH PM90PRO MONITOR					

171	PADS FOR ELECTRICAL STIMULATOR (ENDOMED 182)					
172	EEG PASTE TEN 20					
173	SPOT MONITOR WELCH ALLYN MODULAR CSM TEMPERATURE PROBE					
174	PENNARD STETHOSCOPE					
175	VENTILATOR COMPRESSOR					
176	Portable Suction Machine					
177	Suction Jar with Stand					
178	Pressure Transducer PICCO					
179	DRY Batteries	12V 24AH				
		12V 12AH				
		12V 65AH				
		12V 85AH				
		12V 2.3AH				
		12V 9AH				
		12V 07AH				
180	Ambo Bag Adult (Autoclavable)					
181	Ambo Bag Peads (Autoclavable)					
182	Video Larynscope					
183	Heating element 6000watt with seal					
184	Hearing AID					
185	Infantometer					
186	Stadiometer					
187	CENTRIFUGE BLOOD BAG REFRIGERATED (12 BAGS)					
188	Defibrillator					
189	ECG with caring Trolley					
190	Binocular Microscope Standard Set					
191	Battery 11.1 Volt, 4500 mAH					

9. Evaluation Criteria

No chance will be provided for re-submission of secondary documentation. The bidders must carefully read the instructions; Non-compliance to the stated instruction may lead to their technical disqualification.

(Technical Evaluation Marks: 50)

S.no	Criteria	Details	Marks
1	Firm Experience	05Years of relevant experiences in Hospital Supplies: <ul style="list-style-type: none"> Attached copies of Purchase Orders/Work Orders/Contract Agreement. 	15
2	Sample	<ul style="list-style-type: none"> Excellent Good Quality Poor Quality Poor Sample Should lead to disqualified (non-responsive)	25 15 0
3	Bank Statements	Last three (03) years bank statements (IT -1, IT – 2)	09
4	Audit Report	Audit report of last three (03 years)	09
	Supplier's Warehouse	Warehouse at Peshawar	06
		Warehouse at national level	03
5	Affidavit	The bidder shall provide an undertaking on Stamp Paper that the bidder has not been declared black listed by any Governmental/ Semi-Governmental institutions.	Mandatory
6	Brand Must be Mentioned with all Quoted Items		Mandatory
7	Registration with Govt: Authorities	<ul style="list-style-type: none"> Income Tax/Sales Tax NTN 	Mandatory
		<ul style="list-style-type: none"> Professional Tax/ KNTN 	3
		Total Marks	70

Financial Criteria (30 Marks):

S #	Parameters	Sub-Parameters	Total Marks:
	Price		30
		Lowest Price will get full marks. The formula to calculate the marks for the price submitted is: [Lowest Price (Fm)÷Price of Bid under consideration (F)] x100 x 0.30	30

7. Award of Contract:

- a. After opening of the bids of the firm / contractor quoting the lowest rates in each item are accepted.

8. Payment:

- a. No advance payment will be permissible.
- b. The payment will be made after successful supply and inspection.





LADY READING HOSPITAL
MEDICAL TEACHING INSTITUTION PESHAWAR

TENDER NOTICE 2024-25

Sealed tenders are invited for the supply of following items for LRH Peshawar. Tender should reach to the office of the undersigned on or before **-2024 at 11.00 AM** which will be opened on the same day at **11:30 AM** in presence of bidders or their representatives. The Bid security shall be submitted from the account of the firm/bidder/contractor who submits the bid in favor of Hospital Director, Lady Reading Hospital Medical Teaching Institution Peshawar.

S. No	Name of Item	Required Bid Security	Procurement Method
1.	Medical Non-Medical & Accessories	Rs. 200,000/-	Single Stage Two Envelop

TERMS AND CONDITIONS: -

1. Only typed tender on original prescribing letter pad, sealed & signed should be submitted, the quoted Price must be printed and hand written quoted price is not be acceptable. The tenders must be according to hospital specification; **double rates** for one item will not be accepted.
2. No conditional tender will be acceptable.
3. Pre-bid Meeting if anyone want to attend/participate will be held on **- -2024 at 11:00** Am in the MMD Committee Room of the LRH.
4. A detailed list of specification of required items can be obtained from the Material Management Department of LRH during working hours (**Saturday is not working day**). The specification is also available on our web-site www.lrh.edu.pk
5. The bid security in the shape of Pay order (PO) will not be accepted.
6. The competent Authority has the right to reject all bids under Rule 47 of the Khyber Pakhtunkhwa Public Procurement Regulatory Authority (KPPRA) Procurement Rules 2014, after assigning valid reasons.

NOTE: All the bidders are directed to submit their samples for quoted items with central receiving point (CRP) MMD LRH-MTI, for sample evaluation within 10 days after Tender opening. After the due date samples will not be accepted.

Associate Hospital Director
Lady Reading Hospital, Peshawar
Medical Teaching Institute

Hospital Director
Lady Reading Hospital, Peshawar
Medical Teaching Institute

Address of Bid Submission/Opening:

Lady Reading Hospital/MTI Peshawar, Khyber Pakhtunkhwa,

Telephone(s) : 0092-91-9211927 Ext:1702